

Features and benefits

- Excellent dispensing accuracy with $\pm 0.5\%$
- 100 % traceability with laser-etched part-number, lot number and use-by date
- Eight different sizes for filling between 0.0034 fl oz (0.1 ml) to more than 8.45 fl oz (250 ml)
- Full biopharmaceutical compliance with USP <87>, Ph. Eur. 6.8 Chapter 3.1.9, ISO 10993 and FDA CFR 177.2600



Technical specifications

	Accusil
Certification and compliance	FDA 21CFR177.2600, ISO 10993, Ph. Eur. 6.8 3.1.9, USP <788>, USP <85>, USP <87>
Sterilisation methods	Autoclave, EtO, Gamma, X-ray
Autoclave stability	121 °C 30 minutes
Gamma stability	Max 50kGy
Length	10 - 152 m
Bore size	0.5 - 8 mm
Bore size	0.02 - 0.315 in
Wall thickness	1.6 - 2.2 mm
Wall thickness	0.063 - 0.087 in
Colour	Colorless
Transparency	Translucent
Spallation	Low
Operating temperature range	-20 °C to 80 °C
Operating temperature range	-4 °F to 176 °F
Hardness, shore A (5sec)	58 to 64
Specific gravity	1.15 g/ml
Compression set	$\leq 30\%$
Shelf life	5 years

See Validation guide for more tubing certification and compliance details

Materials of construction

	Accusil
Tubing	Platinum-cured silicone

Product Codes

Dispensing volume in ml*	Bore		Wall		Coil	
	mm	inch	mm	inch	10m/32.8ft	150m/492ft
< 0.5	0.5	1/50	1.6	1/16	84-103-005	84-104-005
0.5 - 1.0	0.8	1/32	1.6	1/16	84-103-008	84-104-008
1.0 - 1.7	1.2	3/64	1.6	1/16	84-103-012	84-104-012
1.7 - 7.0	1.6	1/16	1.6	1/16	84-103-016	84-104-016
7.0 - 12.0	3.2	1/8	1.8	1/14	84-103-032	84-104-032
						125m/410ft
12.0 - 22.0	4.8	3/16	2.0	0.0787	84-103-048	84-104-048
						90m/290ft
22.0 - 35.0	6.0	6/25	2.1	0.0827	84-103-060	84-104-060
						65m/213ft
> 35.0	8.0	5/16	2.2	0.0866	84-103-080	84-104-080

* (For optimum dispensing accuracy and speed. Based on dispensing of viscosity like water)

Disclaimer: The information contained in this document is believed to be correct but Watson-Marlow Limited accepts no liability for any errors it contains and reserves the right to alter specifications without notice. It is the users responsibility to ensure product suitability for use within their application. Watson-Marlow and WMArchitect are registered trademarks of Watson-Marlow Limited. Bio Y, BioClamp, BioBarb, FlatBioEndCap, BioEndCap, BioValve and BioTube applicator are Trademarks of BioPure Technology Limited. Tri-Clamp is a registered trademark of Alfa Laval Corporate AB. Q-Clamp is used under license.



wmfts.com/global
31 March 2025